


# United Kingdom Accreditation Service

---

---

## ACCREDITATION CERTIFICATE



**TESTING LABORATORY  
No. 2604**

### **Enquin Environmental Limited**

is accredited in accordance with the recognised International Standard ISO/IEC 17025:2005  
General Requirements for the competence of testing and calibration laboratories.

This accreditation demonstrates technical competence for a defined scope as detailed in and at the locations specified in the schedule to this certificate, and the operation of a laboratory quality management system (refer joint ISO-ILAC-IAF Communiqué dated January 2009).

The schedule to this certificate is an essential accreditation document and from time to time may be revised and reissued by the United Kingdom Accreditation Service. The most recent issue of the schedule of accreditation, which bears the same accreditation number as this certificate, is available from the UKAS website [www.ukas.com](http://www.ukas.com).

This accreditation is subject to continuing conformity with United Kingdom Accreditation Service requirements. The absence of a schedule on the UKAS website indicates that the accreditation is no longer in force.



*Accreditation Manager, United Kingdom Accreditation Service*

**Initial Accreditation date  
3 November 2003**

**This certificate issued on  
03 October 2012**


UKAS is appointed as the sole national accreditation body for the UK by The Accreditation Regulations 2009 (SI No 3155/2009) and operates under a Memorandum of Understanding (MoU) with the Department for Business, Innovation and Skills (BIS).

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>2604</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Enquin Environmental Limited</b>	
	Issue No: 011    Issue date: 27 July 2018	
	<b>1 Ringside Business Centre</b> Heol Y Rhosog Wentloog Cardiff CF3 2EW	<b>Contact: Ben Holderness</b> Tel: +44 (0)29 2053 5090 Fax: +44 (0)29 2077 8633 E-Mail: <a href="mailto:bholderness@enquin.co.uk">bholderness@enquin.co.uk</a> Website: <a href="http://www.enquin.co.uk">www.enquin.co.uk</a>
Testing performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> 1 Ringside Business Centre Heol Y Rhosog Wentloog Cardiff CF3 2EW  <b>Local contact</b> Ben Holderness Tel: +44 (0)29 2053 5090 Fax: +44 (0)29 2077 8633 Email: <a href="mailto:asbestos@enquin.co.uk">asbestos@enquin.co.uk</a> Website: <a href="http://www.enquin.co.uk">www.enquin.co.uk</a>	Health and Hygiene	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



2604  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Enquin Environmental Limited**  
Issue No: 011 Issue date: 27 July 2018

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 Membrane Filter Method using Phase Contrast Microscopy (PCM)	A, B
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure in the Procedures Manual,Section 2)	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure in the Procedures Manual, Section 2)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure, Procedures Manual, Section 3, using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for asbestos identification	Documented In-House Method – Procedures Manual, Section 4, based upon HSG 248 and HSE HSG264	B
END			